

File No. 10-14/CHEM/CFSL/2018/GP

**Central Forensic Science Laboratory
Central Bureau of Investigation
Block No.4, CGO Complex, Lodhi Road,
New Delhi - 110 003**

Corrigendum in respect of advertisement uploaded on CBI website on 20.04.2018 for purchase of gas purification panels, control boxes, manifold system, cylinder holding brackets and other accessories for nitrogen, zero air, hydrogen and helium gas cylinders and its installation.

Based on the queries received from the prospective bidders in reference to the above advertisement following clarification offered/amendments are made in the technical specifications:

1. The stainless steel items of AISI 316 grade to be provided.
2. Suitable Flash Back Arrestor for hydrogen gas should be provided.
3. Micron filters for capturing foreign particles for every gas to be provided.

Interested firms/vendors may send their sealed quotations to "The Director, Central Forensic Science Laboratory, CBI, Block No.-4, CGO Complex, Lodhi Road, New Delhi -110003" **within 20 days** from the posting of this corrigendum on website.

The firms who have sent the quotations earlier may send their revised quotations incorporating the above amendments if these are not included in there earlier quotations.

Other terms and conditions remains the same.



**(N. Ravi)
I/C Director**

CFSL (CBI), New Delhi
Tel: (011)-2436 1396
Fax: 011-24360742