

File No: 2-4 / 2016- CFSL (Bio) / 1728
Central Forensic Science Laboratory (CBI),
Block No:4, CGO Cómplex,
Lodhi Road, New Delhi

Dtd: 02/05/2016

To
The Senior System Analyst,
CBI Comand Centre, 6th floor, CBI HQRS,
Plot No: 5B, CGO Complex, Lodhi Road, New Delhi.

Sub: Publication in CFSL / CBI website reg;


Sir,

CFSL (CBI), New Delhi is in the process of procuring Autosomal DNA STR amplification kits And Y Chromosome DNA STR Amplification Kit for it's use in the DNA Profiling laboratory . Necessary arrangements may kindly be made for uploading the bellow mentioned format in the CFSL / CBI website for a period of 21 days for inviting necessary quotations.

Financial Quotations along with required booklet / literature are required from interested suppliers / manufacturers for supply of Autosomal DNA STR amplification kits and Y chromosome DNA STR Amplification kit of the following specifications and quantity. The quotations addressed to The Director, CFSL (CBI), New Delhi may be submitted in the office of the Director, Central Forensic Science Laboratory (CBI), Block -4, CGO Complex, Lodhi Road, New Delhi within 21 days of publication of this notice.

Name of the item	Quantity required	Specification
Autosomal DNA STR amplification kit	200 reactions	<ol style="list-style-type: none">To be able to amplify at least 16 autosomal human identification DNA markers including amelogenin marker.It must include the 13 CODIS loci.It should be compatible with genemapper ID /IDX soft ware for data analysis.It should be compatible with 3100 /3130XL genetic analyzer of Invitrogen life technology for analysis.Desired bin set, or any other analytical soft ware etc if required is to be provided by the supplier free of cost and installed in the genetic analyzer.Amplification kits should be validated for forensic applications.Application support is to be provided by the supplier as and when desired.
Y Chromosome DNA STR amplification kit	100 actions	<ol style="list-style-type: none">To be able to amplify at least 17 Y Chromosome Human Identification DNA Markers..It should be compatible with genemapper ID /IDX soft ware for data analysis.It should be compatible with 3100 /3130XL genetic analyzer of Invitrogen life technology for analysis.Desired bin set, or any other analytical soft ware etc if required is to be provided by the supplier free of cost and installed in the genetic analyzer.Amplification kits should be validated for forensic applications.Application support is to be provided by the supplier as and when desired.

1045/50
02/05/16


Dr Rajinder Singh
Director, CFSL (CBI), New Delhi